

**BLUE NOTICE** Please read, sign, & follow any additional directions below:

To: RCS Staff  
From: Laura Martin  
Date: May 2017

Re: Documentation to eliminate potential for accidental loss of Nitric Oxide

Notes: Binder with NO delivery documentation will remain in the NICU RCS equipment room.  
Contact Supervisor as soon as NO is initiated or discontinued. Return NO document to binder

Instructions: Purge/Performance, Calibration and Training are only to be done with Calibration cylinder clearly marked with red flag.

SD  
MB  
algridB  
MB  
RB  
GB  
RC  
W  
SC  
J  
BAT  
MD  
RE  
GRO  
RF  
MDF  
R  
HH  
KH

Ang, Steven  
Bandian, Leo  
Barnoski, Ingrid  
Bell, Malika  
Booze, Joseph  
Burns, Rachael  
Burns, Greg  
Calma, Franco  
Carothers, Logan  
Castillo, Sam  
Cespedes, Clara  
Chow, Joanna  
Decuire, Byron  
Diaz, Martha  
Ennis, Robert  
Fazel Darbandi, Jill  
Fejerang, Rosemarie  
Ferrer, Jason  
Fisher, Paul  
Fong, Myron  
Foster, Loretta  
Gatchalian, Leslie  
Hammoudeh, Hanaa  
Ho, Pik Wah (Kelly)

JH  
JJ  
K  
K  
KW  
SM  
EM  
LM  
AM  
CM  
CM  
JN  
JN  
MN  
H  
LP  
RP  
AM  
AL  
S

Hsieh, Jenny  
Johal, Jasmeet  
Kallet, Richard  
Kelly, John  
Lear, Veronica  
Leonor, Nicholas  
Lewis, Eula  
Mahboobi, Safi  
Mangalindan, Earl  
Martin, Laura  
Matsushima, Aya  
Milleza, Consuelo  
Mojaddedi, Moman  
Moon, Cheryl  
Nastor, Jessica  
Navarro, Anngie  
Nelson, Mike  
Ong, Hannah  
Pangilinan, Lance  
Perez, Ricardo  
Phillips, Justin  
Quong, Andrew  
Rahming, Chanera  
Saladze, Dimitri

**BLUE NOTICE** Please read, sign, & follow any additional directions below:

---

To: RCS Staff

From: Laura Martin

Date: May 2017

Re: Documentation to eliminate potential for accidental loss of Nitric Oxide

Notes: Binder with NO delivery documentation will remain in the NICU RCS equipment room.  
Contact Supervisor as soon as NO is initiated or discontinued. Return NO document to binder

Instructions: Purge/Performance, Calibration and Training are only to be done with Calibration cylinder clearly marked with red flag.

ES

Saliba, Edward

LS

Schletz, Susan

MS

Scholz, Misty

AB

Sikhounchanh, Annie

TS

Summers, Travis

JS

Sylvester, Josephine

N

Vides, Andrew

Vinson, David

SW

Waugh, Shane

GW

Wedeking, Grey

DW

Whiting, Douglas

WY

Yip, Vivian

**INOMaxDsir Plus Delivery System Check List**

0000000000  
**BABY JANE**  
 Patient MRN & Name Stamp

RCS Staff Name and LCR (079): LAURA INGALS WILDER

	<u>Date:</u>	<u>Time:</u>
<b>Therapy initiated:</b>	5-18-17	5:00 AM
<b>Supervisor Observing:</b>	JOHN SCHINDLER	
<b>Patient Name and MRN: ENTERED INTO INOMAX</b>	YES	
<b>RCS AOD NOTIFIED:</b>	Notification by email is appropriate: YES <input checked="" type="checkbox"/>	
	<u>Date:</u>	<u>Time:</u>
<b>Initial systems test completed successfully: <u>Testing performed with Calibration tank only!</u></b>	5-18-17	4:00 AM

	<u>Date:</u>	<u>Time:</u>
<b>Nitric Oxide Discontinued:</b>	5-19-17	5:00 AM
<b>Nitric Oxide Tanks turned off:</b>	YES	
<b>Regulator removed:</b>	YES	
<b>Red lock, locked in place:</b>	YES	
	<u>Date:</u>	<u>Time:</u>
<b>Supervisor notified NO discontinued:</b>	5-19-17	5:00 AM
<b>Data loaded onto Data Collection Mode: to be completed by Equipment Manager or Supervisor</b>	YES	

RCS Staff Discontinuing Nitric Oxide Signature and LCR (079):  
 Name/079 LAURA INGALS WILDER